VU Research Portal

Combined macro- and microrheometer for use with Langmuir monolayers

Walder, R.; Schmidt, C.; Dennin, M.

published in

Review of Scientific Instruments
2008

DOI (link to publisher)

[10.1063/1.2937649](https://doi.org/10.1063/1.2937649)

document version

Publisher's PDF, also known as Version of record

[Link to publication in VU Research Portal](#)

citation for published version (APA)

Walder, R., Schmidt, C., & Dennin, M. (2008). Combined macro- and microrheometer for use with Langmuir monolayers. *Review of Scientific Instruments*, 79(6), 063905. <https://doi.org/10.1063/1.2937649>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

E-mail address:

vuresearchportal.ub@vu.nl

Combined macro- and microrheometer for use with Langmuir monolayers

Robert Walder,¹ Christoph F. Schmidt,² and Michael Dennin¹

¹*Department of Physics and Astronomy, University of California, Irvine, California 92697, USA*

²*Drittes Physikalisches Institut, Fakultät für Physik, Georg-August-Universität, 37077 Göttingen, Germany*

(Received 11 January 2008; accepted 8 May 2008; published online 16 June 2008)

A Langmuir monolayer trough that is equipped for simultaneous microrheology and standard rheology measurements has been constructed. The central elements are the trough itself with a full range of optical tools accessing the air-water interface from below the trough and a portable knife-edge torsion pendulum that can access the interface from above. The ability to simultaneously measure the mechanical response of Langmuir monolayers on very different length scales is an important step for our understanding of the mechanical response of two-dimensional viscoelastic networks. © 2008 American Institute of Physics. [DOI: 10.1063/1.2937649]

Soft-matter systems play critical roles both in technological applications and in biology. Most examples of soft matter are complex fluids, exhibiting both solid- and fluidlike responses. One of the challenges in the soft-matter research is the quantitative understanding of such behavior. Complex dynamics are typically the result of complex structure spanning a range of length scales. Often, the large-scale response of the system depends in a nontrivial fashion on the interaction between mesoscopic structures in the system. The local response within one of these mesoscopic structures can be very different from the macroscopic response. For example, aqueous foam is composed of gas bubbles separated by liquid films. Measurements within a liquid film suggest the material is fluid. However, for applied stress below the yield stress, the macroscopic foam acts as a solid. Due to various additives, the liquid film itself may be viscoelastic and the details of this “microscopic” response can, in turn, directly impact the details of the macroscopic behavior.

A system of great biological interest, for which length scales play an important role are semiflexible polymer networks. Many structures in biology are composed of semiflexible networks of protein filaments, such as the actin cytoskeleton, that are also connected with lipid membranes. The need to understand both the microscopic and macroscopic responses of such a composite system provided much of the motivation for the apparatus that we report on in this paper. The other motivation was the need to make measurements on a robust model system. For this, we have chosen to focus on Langmuir monolayers as a model for lipid structures. Langmuir monolayers are single-molecule-thick layers of amphiphilic molecules at the air-water interface, and thereby resemble a half of a lipid bilayer which forms cell membranes. There exists a well-developed set of tools for studying their mechanical properties. This offers many advantages when pursuing mechanical studies and it allows us to build on existing technology.

One can categorize the existing methods for measuring the viscoelastic properties of Langmuir monolayers according to the length scale probed by the technique. The more common techniques that probe the system on relatively large length scales include needle viscometers,¹ channel viscometers,^{2–5} the knife-edge torsion pendulum,^{6–10} and

light scattering.^{11,12} Our group has also developed a Couette-style rheometer that expands on standard knife-edge torsion pendulum techniques.¹³

The first use of microscopic-scale local techniques in Langmuir monolayers was based on measuring relations between translational and rotational mobilities of particles dif-

FIG. 1. (Color online) Two pictures of the apparatus. The top image shows the view from where the laser enters the optics below the trough. The laser enters from the lower left side of the image. The bottom image is the view from the opposite side of the trough. The black case in the background houses the laser. (For a scale, the spacing between the holes on the optical table are 1 in.)

FIG. 2. (Color online) Entire experimental setup.

fusing in the monolayer.^{14,15} In this regard, a useful feature of Langmuir monolayers is their rich phase behavior, with the equivalent of gas, liquid, liquid crystal (often referred to as liquid condensed), and crystalline phases. Because they often exhibit phase coexistence, the domains of one phase can be used as the “particles” to probe the properties of the other phase. For example, experiments have included measuring the diffusion of liquid crystalline domains¹⁶ and using liquid crystalline domains as particles for measuring velocity profiles.⁵ More recently, Brewster-angle microscopy has been used to study the motion of domains.^{17,19} Our apparatus uses microrheological techniques that rely on the introduction of tracer particles. It is based on similar techniques discussed in Refs. 18–21. However, as discussed in more detail later, we expand on these techniques in several ways.

In considering the design of an apparatus that can measure rheological properties on multiple length scales, we decided to pursue a modular design that combines two established rheological techniques: particle microrheology and the knife-edge torsion pendulum. Both of these measure the complex shear modulus G . In linear response, the complex shear modulus provides the connection between an oscillatory strain $\gamma(\omega)$ and stress $\sigma(\omega)$ as a function of the oscillation frequency ω through the relation: $\sigma(\omega) = G\gamma(\omega)$. The real part of G is the elastic modulus and the imaginary part is related to the viscous response (i.e., energy dissipation in the system). The difference between the microrheology measurements and the knife-edge torsion pendulum is in the length scales being probed. Typically, microrheology uses micron-size beads embedded in the material and the torsion pendulum is on the order of millimeter to centimeter. We will briefly review each technique and discuss the design goals

for each of these. Then, we will describe our apparatus in detail.

The most basic form of microrheology uses thermal position fluctuations of tracer particles (typically micron-sized spheres) embedded in the material as probes.^{22–24} The displacement of a single spherical probe in response to a force on it is described by a response function which is given by the generalized Stokes–Einstein relation:^{23,25} (GSER)

$$u(\omega) = \alpha(\omega)f(\omega) = \frac{f(\omega)}{6\pi aG(\omega)},$$

where $\alpha(\omega)$ is the complex response function and $G(\omega)$ is the complex shear modulus. For purely viscous samples, $G(\omega) = -i\omega\eta$. For a thermally fluctuating particle, one can calculate the power spectral density of the position fluctuations, which by the fluctuation-dissipation theorem, is related to the imaginary part of the response function.

This method is termed passive microrheology because nothing but random thermal forces experienced by the particles are used as the driving force. Our system is based on a procedure for determining the shear moduli that is fully discussed in Schnurr *et al.*²² Essentially, one uses the fluctuation-dissipation theorem to relate the power spectral density of the thermal displacements of the bead to the imaginary part of the complex response function $\alpha''(f)$. When measuring $\alpha''(f)$ over a large enough frequency range, a Kramers–Kronig integral can be used to recover the real part of the response function $\alpha'(f)$, and the total complex response function $\alpha(f)$ is then related to the complex shear modulus $G(f) = G'(f) + iG''(f)$ by the GSER

FIG. 3. Isotherm performed on the DPPC monolayer at room temperature.

$$G(f) = \frac{1}{6\pi a \alpha(f)}.$$

Therefore, a critical design element in the system is the ability to measure over a wide range of frequencies. This will be discussed in the detailed description of the apparatus. The final key experimental component for microrheology is the use of optical tweezers to trap the probe particle.²⁶

The knife-edge torsion pendulum method is well established to probe viscoelastic properties of surface layers on air-water interfaces. Essentially, one suspends a Teflon disk

with a knife edge at its circumference in the monolayer. The disk is driven with a known applied torque, which corresponds to an applied stress for the monolayer. The angular response of the disk corresponds to the strain response of the monolayer. One must correctly account for the contribution of the fluid subphase and the details of this are given in Ref. 10. The key elements in this case are the driving mechanism and the measurement of the rotation angle. For the application discussed here, it was necessary to adopt a design that could be easily integrated with the microrheology technique.

The integration of the knife-edge torsion pendulum with the microrheology technique required the design of a new Langmuir monolayer trough with full optical access from below and the adaptation of our existing knife-edge torsion pendulum. Access from below was necessary for the optical tweezers and particle tracking in order to harness the full power of microrheology. In addition, for added flexibility, we provided access from below for a fluorescence microscope so as to leave space above the monolayer for the knife-edge torsion pendulum. The main mechanical adaptation necessary for the knife-edge torsion pendulum was the design of a modular unit that can function as a stand-alone device. This makes it possible to remove the device for the use of additional tools in combination with the optical tweezers. It has the added benefit that the knife-edge torsion pendulum can be used and calibrated in a separate mini-Couette rheometer. We will first describe the Langmuir trough and integrated microrheology apparatus and then we will separately discuss the design of the stand-alone knife-edge torsion pendulum. The combination of the two techniques into one instrument creates a powerful, high-precision instrument capable of ex-

FIG. 4. (Color online) Fluorescence microscope.

FIG. 5. (Color online) Optical tweezer.

ploring mechanical properties of surface films that were previously not accessible.

The integrated Langmuir trough consists of three optical instruments and a surface tensiometer. Figure 1 provides two views of the apparatus and Fig. 2 is a schematic of the overall apparatus. Details of each portion of the apparatus are provided in the following sections. The integration of these systems into one master instrument allows us to simultaneously image and mechanically manipulate surface films in the Langmuir trough and track embedded particles. A number of the components are standard equipment for Langmuir monolayer troughs. For example, a commercial surface pressure sensor²⁷ is used to characterize the phase of the monolayer. A typical isotherm is illustrated in Fig. 3 for a dipalmitoylphosphatidylcholine (DPPC) monolayer. The isotherm shows small steps which come from the motor action in compressing the monolayer. The motor compresses the monolayer approximately 0.5 cm over 1 s and then allows 4 s for relaxation. The sampling rate for all data is 4 Hz. In addition, the isotherm exhibits a slightly lower area per molecule at the transition to the condensed phase than one finds for a typical DPPC isotherm. The major cause of this effect was the addition of the probe particles but there was probably some contribution from errors in solution concentration generated by the age of the solution. Since this discrepancy was not relevant to the present experiments, we did not correct for the various effects.

Due to space and cost constraints, we have chosen to use a custom built fluorescence microscope that images the monolayer from below instead of mounting the system on a full commercial fluorescence microscope (see Fig. 4). Fluorescence

microscopy is essential to be able to specifically detect molecules at the interface that cannot be detected using standard bright-field microscopy. Selected populations of molecules in the sample (in our case lipids and/or actin) are modified by chemically attaching a fluorophore.

In our optical configuration, we use a mercury arc lamp²⁸ combined with appropriate excitation filters mounted in a filter wheel, to excite fluorescence, and the corresponding emission filters on a filter wheel connected to our intensified charge coupled device (CCD) camera²⁹ to detect fluorescence. This configuration allows us to use different fluorophores during our experiments and thus allows us to selectively image different components of our sample. The two filter cubes we have currently installed contains fluorescein isothiocyanate (FITC) and rhodamine filters. One particularly useful application of this capability is to use one fluorophore for the lipid monolayer (such as Bodipy) and to use another fluorophore for proteins introduced into the sub-phase (such as rhodamine actin).

An optical tweezer essentially consists of a laser focused to a diffraction limited focus through a high-numerical-aperture objective lens. The light intensity gradient creates a force on objects of a higher index of refraction than the surroundings. The two basic components of an optical tweezers are thus a laser³⁰ and an objective lens³¹ (see Fig. 5). The objective lens needs to have a high numerical aperture to create a steep enough intensity gradient.³² The water immersion objective lens we use for our optical tweezer system has a numerical aperture of 1.0 and a working distance of 1.5 mm. The Langmuir trough is mounted on a three-axis translation stage, while the objective lens is fixed in position.

FIG. 6. Imaging a trapped bead in a DPPC monolayer at a surface pressure of 6 mN/m.

The laser is then focused onto the surface by lowering the Langmuir trough using a micrometer that controls the height of the trough relative to the objective lens. If water evaporates during the experiment, the Langmuir trough height is adjusted to refocus to the surface. The long working distance of the objective lens prevents a meniscus from forming at the surface.

Additional optical components were added to the laser/objective lens system in order to control trap strength and to allow us to steer the trap location in the sample. An optical isolator prevents backscattering which can cause laser instability. A beam expander increases the beam diameter fourfold

in order to use the maximal numerical aperture by filling the back aperture of the objective lens. Two 100 mm focal length lenses form a telescope for beam steering. The first lens is mounted on a three-axis translation stage. Translation of this lens steers the trap location in three dimensions within a range of several microns in each axis. An example of an image of a particle in the trap (using the fluorescence microscope) is shown in Fig. 6.

A key element of the microrheology system is the particle displacement measurement. The thermal position fluctuations are close to and below the limits of the spatial optical resolution that can be achieved with standard video particle tracking. In addition, the frame rate of the CCD camera used for fluorescence microscopy is limited to a range of 10–100 Hz. This imposes a limit on the range of frequencies over which the complex shear modulus can be determined by video particle tracking. Using the quadrant diode based displacement detection system described in the next section (see Fig. 7), we can probe the frequency response of the system up to 100 kHz.

The backscattered laser light from the trapped particle is imaged onto a quadrant photodiode in a method similar to that described by Helfer *et al.*³³ When forward-scattered light can be collected, one can use back-focal plane interferometry.^{34,35} In our case, we have no optical access from above and therefore use backscattered light. The detection optics consists of a beam splitter, imaging lenses, a 45° 1064 nm laser line mirror, and a quadrant photodiode.³⁶ The beam splitter is used to reflect some of the backscattered laser light from the trapped particles into the imaging optics. The lenses are positioned to form an image of the particle on

FIG. 7. (Color online) Particle tracking.

FIG. 8. Power spectral density of a DPPC monolayer at a surface pressure of 6 mN/m. The data have been smoothed by logarithmic averaging.

the quadrant photodiode with a final magnification of $100\times$. The laser line mirror is used both as a filter to prevent any visible light from the fluorescence imaging from reaching the quadrant photodiode and to center the backscattered light onto the quadrant photodiode. In addition, the laser line mirror protects the CCD camera from being damaged by the backscattered laser light. The quadrant photodiode then records the position of the particle with high spatial and temporal resolution in terms of the four output photocurrents. The sampling frequency is typically 66 kHz and the number of points sampled is 2^n , in order to use fast Fourier transform for data analysis.

The use of backscattered laser light means that we are photon limited and that our measurements cannot quite achieve the performance of interferometric based systems.^{35,37} However, this optical system allows us to expand previous measurements of static viscosity and drag coefficient measurements for monolayers^{19,21} into a significantly larger frequency range. A set of data is shown in Figs. 8–10 for the power spectral density, elastic modulus, and

FIG. 9. Elastic modulus of a DPPC monolayer at a surface pressure of 6 mN/m, calculated from the power spectral density shown in Fig. 8.

FIG. 10. Loss modulus of a DPPC monolayer at a surface pressure of 6 mN/m. This has been calculated from the power spectral density shown in Fig. 8.

loss modulus, respectively, of a $1\text{ }\mu\text{m}$ diameter bead in a DPPC monolayer at a surface pressure of 6 mN/m and an area per molecule of $55\text{ }\text{\AA}^2$ per molecule. In principle, our instrument can measure the complex shear modulus over the full range of surface pressure. The main limitation to the technique is flow in the monolayer (due to convection, air currents, etc.). If the flow is strong enough, it can prevent trapping of the beads. However, we are able to limit the flow with two different methods. First, the use of a cover eliminates air currents. Second, a small ring can be placed in the subphase just below the surface of the water. Once the phase of the monolayer is selected by the compression of the barriers, the surface of the water can be lowered and a small section of the monolayer can be isolated by the ring. This very efficiently limits flow in the subphase. We are currently working on a systematic study of DPPC monolayers in all phases, which is the subject of a future publication.

Because we have designed the main components for microrheology to be located below the trough, we can access the surface of the water with a standard torsion, knife-edge rheometer. For this, we have constructed a removable device. Figure 11 shows a schematic. The key elements are the suspension of a circular Teflon disk by a torsion wire. The support for the disk also contains a magnet coil (green rectangle in Fig. 11). This coil is situated within a larger coil that generates a magnetic field, indicated by the x 's in Fig. 11. Additionally, there is a fixed magnet attached to the coil on the disk support.

The coils serve a dual purpose. With the coils, we are able to both measure the angular position of the disk and provide an external torque to the disk to drive it at a defined frequency. The details of measuring the complex shear modulus with this method are provided in Ref. 10. The coils and torsion pendulum match the specifications reported in Ref. 13. We have modified the support structure to allow for portability of the device.

We acknowledge the support of NSF-DMR-0354113. We also thank Thomas Fischer for helpful discussions. R. Walder acknowledges support through a travel fellowship

FIG. 11. (Color online) Schematic of the portable torsion pendulum.

from the Institute for Complex Adaptive Matter. C. F. Schmidt acknowledges support by the DFG/SFB755 Nano-scale Photonic Imaging and by the DFG Center for Molecular Physiology of the Brain (CMPB).

- ¹C. F. Brooks, G. G. Fuller, C. W. Frank, and C. R. Robertson, *Langmuir* **15**, 2450 (1999).
- ²W. D. Harkins and J. G. Kirkwood, *J. Chem. Phys.* **6**, 53 (1938).
- ³L. E. Copeland, W. D. Harkins, and G. E. Boyd, *J. Chem. Phys.* **10**, 357 (1942).
- ⁴M. Sacchetti, H. Yu, and G. Zografi, *Rev. Sci. Instrum.* **64**, 1941 (1993).
- ⁵D. K. Schwartz, C. M. Knobler, and R. Bruinsma, *Phys. Rev. Lett.* **73**, 2841 (1994).
- ⁶R. J. Mannheimer and R. S. Schechter, *J. Colloid Interface Sci.* **32**, 195 (1970).
- ⁷S. S. Feng, R. C. MacDonald, and B. M. Abraham, *Langmuir* **7**, 572 (1992).
- ⁸B. M. Abraham, K. Miyano, S. Q. Xu, and J. B. Ketterson, *Rev. Sci. Instrum.* **54**, 213 (1983).

- ⁹J. Krgel, S. Siegel, R. Miller, M. Born, and K.-H. Schano, *Colloids Surf., A* **91**, 169 (1994).
- ¹⁰R. S. Ghaskadvi, J. B. Ketterson, R. C. MacDonald, and P. Dutta, *Rev. Sci. Instrum.* **68**, 1792 (1997).
- ¹¹S. Hård and H. Löfgren, *J. Colloid Interface Sci.* **60**, 529 (1976).
- ¹²D. Langevin, *J. Colloid Interface Sci.* **80**, 412 (1981).
- ¹³R. S. Ghaskadvi and M. Dennin, *Rev. Sci. Instrum.* **69**, 3568 (1998).
- ¹⁴B. D. Hughes, B. A. Pailthorpe, L. R. White, and W. H. Sawyer, *Biophys. J.* **37**, 673 (1982).
- ¹⁵P. G. Saffman and M. Delbruck, *Proc. Natl. Acad. Sci. U.S.A.* **72**, 3111 (1975).
- ¹⁶J. F. Klingler and H. M. McConnell, *J. Phys. Chem.* **97**, 6096 (1993).
- ¹⁷A. T. Ivanova, J. Ignes-Mullol, and D. K. Schwartz, *Langmuir* **17**, 3406 (2001).
- ¹⁸S. Wurlitzer, P. Steffen, and Th. M. Fischer, *J. Chem. Phys.* **112**, 5915 (2000).
- ¹⁹P. Steffen, P. Heinig, S. Wurlitzer, Z. Khattari, and Th. M. Fischer, *J. Chem. Phys.* **115**, 994 (2001).
- ²⁰S. Wurlitzer, C. Lautz, M. Liley, C. Duschl, and Th. M. Fischer, *J. Phys. Chem. B* **105**, 182 (2001).
- ²¹M. Sickert and F. Rondelez, *Phys. Rev. Lett.* **90**, 126104 (2003).
- ²²B. Schnurr, F. Gittes, F. C. MacKintosh, and C. F. Schmidt, *Macromolecules* **30**, 7781 (1997).
- ²³F. Gittes, B. Schnurr, P. D. Olmstedt, F. C. MacKintosh, and C. F. Schmidt, *Phys. Rev. Lett.* **79**, 3286 (1997).
- ²⁴F. C. MacKintosh and C. F. Schmidt, *Curr. Opin. Colloid Interface Sci.* **4**, 300 (1999).
- ²⁵A. J. Levine and T. C. Lubensky, *Phys. Rev. Lett.* **85**, 1774 (2000).
- ²⁶A. Ashkin, J. M. Dziedzic, J. E. Bjorkholm, and S. Chu, *Opt. Lett.* **11**, 5 (1986).
- ²⁷Nima Technology, Ltd., Surface Pressure Sensor Type PS4, The Science Park, Coventry, CV4 7EZ, England.
- ²⁸Olympus America Inc., Model 5-UL155, 3500 Corporate Parkway, P.O. Box 610, Center Valley, PA.
- ²⁹Cohu Intensified CCD Camera Model 5515-2001, 3912 Calle Fortunada, San Diego, CA 92123-1827.
- ³⁰Spectra Physics Model BL-106C, 1335 Terra Bella Avenue Mountain View, CA 94039.
- ³¹Olympus America Inc., Model 1-UM575, 3500 Corporate Parkway P.O. Box 610 Center Valley, PA.
- ³²M. J. Lang and S. M. Block, *Am. J. Phys.* **71**, 201 (2003).
- ³³E. Helfer, S. Harlepp, L. Bourdieu, J. Robert, F. C. MacKintosh, and D. Chatenay, *Phys. Rev. Lett.* **85**, 457 (2000).
- ³⁴M. W. Allersma, F. Gittes, M. J. Castro, R. J. Stewart, and C. F. Schmidt, *Biophys. J.* **74**, 1074 (1998).
- ³⁵F. Gittes and C. F. Schmidt, *Opt. Lett.* **23**, 7 (1998).
- ³⁶New Focus Model 2309, 2584 Junction Avenue San Jose, CA 95134.
- ³⁷K. Addas, C. F. Schmidt, and J. Tang, *Phys. Rev. E* **70**, 021503 (2004).