VU Research Portal

Frozen storage of serum does not affect cholesterol and triglyceride concentration in lipoproteins as separated by gradient ultracentrifugation

Tiedink, H.G.M.; Katan, M.B.

published in

Clinical Chemistry
1988

document version

Publisher's PDF, also known as Version of record

[Link to publication in VU Research Portal](#)

citation for published version (APA)

Tiedink, H. G. M., & Katan, M. B. (1988). Frozen storage of serum does not affect cholesterol and triglyceride concentration in lipoproteins as separated by gradient ultracentrifugation. *Clinical Chemistry*, 34, 593-4.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

E-mail address:

vuresearchportal.ub@vu.nl

ity was counted. The figure shows that wells coated with murine (◆—◆) and equine antibody (◇—◇) show a reproducible pattern of reactivity, with an evident increase in counts in the direction away from the identification tab. CVs for counts/min per well within a single strip ranged from 2.1 to 11.9% (murine) and 2.1 to 10.4% (equine), ($n = 8$).

This phenomenon was not seen in a two-site radioassay for α_1 m. Bound antibody was followed by 100 μ L of a 1:640 dilution of human serum. After incubation, wells were washed, an 125 I-labeled mouse anti- α_1 m immunoglobulin of different epitopic specificity was added, and the wells were incubated, washed, and the radioactivity was counted. The increase in counts was absent in the "sandwich" format (■—■); CVs ranged from 2.8 to 5.9% ($n = 8$).

Kricka et al. (1) report an "edge effect," with variability for competition and two-site immunoassays of $\pm 11\%$ and $\pm 18\%$. Researchers should be aware of these various binding phenomena. These results depend upon as-yet-unidentified variables that are associated with individual proteins.

Reference

1. Kricka LJ, Carter TJN, Burt SM, et al. Variability in the adsorption properties of microtitre plates used as solid supports in enzyme immunoassay. *Clin Chem* 1980;26:741-4.

Frozen Storage of Serum Does Not Affect Cholesterol and Triglyceride Concentration in Lipoproteins as Separated by Gradient Ultracentrifugation, Henrica G. M. Tiedink and Martijn B. Katan

(Dept. of Human Nutrition, Agricultural University, De Dreijen 12, 6703 BC Wageningen, The Netherlands)

We investigated the effect of prior storage of serum at -20°C on the concentration of cholesterol and triglycerides measured in lipoprotein fractions separated by density-gradient ultracentrifugation. Fresh, preprandial, normolipemic sera from 14 men and 10 women were divided into five 2.5-mL aliquots (A-E). Aliquots A were stored at 4°C for <24 h. Aliquots B were stored at -20°C for 2.5 h, thawed in air of 20°C , and stored at -20°C for another 18 h. Aliquots C, D, and E were stored at -20°C for 2, 11, and 27 weeks, respectively. After storage, serum lipoproteins were separated by density-gradient ultracentrifugation (1) for 22 h in a Beckman Ti 41 SW rotor at 20°C and 37 000 rpm ($233\,000 \times g$). Lipoproteins were aspirated and densities checked as described (1), and cholesterol and triglycerides

Fig. 1. Cholesterol (upper curves) and triglyceride (lower curves) concentrations measured in lipoprotein fractions after prior storage of sera at -20°C for the times indicated

Time 0 FRESH refers to samples stored at 4°C for less than 24 h; time 0 FROZEN refers to samples that had been stored at -20°C for 2.5 h, thawed, and then stored at -20°C for another 18 h. Symbols: ■, total serum; □, LDL; ▲, VLDL; ○, HDL₂; ●, HDL₃.

were assayed enzymatically (2, 3). The CV for cholesterol was 1.3% at 4.80 mmol/L, 0.9% at 8.96 mmol/L, and for triglycerides 1.6% at 0.91 mmol/L, 1.4% at 1.80 mmol/L. Bias for serum pools provided by the Centers for Disease Control, Atlanta, GA, was -0.7% for cholesterol and $+0.9\%$ for triglycerides. The mean percentage recoveries ranged between 97.3% and 101.2% for cholesterol and between 71.1% and 76.0% for triglycerides. The low analytical recovery of triglycerides was probably caused by small amounts of VLDL sticking to the wall of the tubing used in aspirating the fractions. Comparison of aliquots A and B (Figure 1) showed that freezing per se had no effect on the measured concentration of cholesterol and triglycerides in the various lipoproteins; changes averaged 0.00 mmol/L (range, -0.02 to 0.02 mmol/L). The measured concentration of total cholesterol in serum was 0.09 mmol/L (-1.8%) lower in aliquots B than in aliquots A.

The concentration of cholesterol in the major cholesterol fractions (LDL, HDL₂, and HDL₃) changed on average by less than 4.1% over the full period of 27 weeks (Figure 1). The percentage changes in the minor fractions were sometimes considerably larger, but the absolute changes were similar. There were no systematic up- or downward trend with time in any of the lipoprotein classes. The percentage changes in the triglyceride concentrations were $<5\%$ (Figure 1) for the major triglyceride fractions (VLDL and LDL) during the first 11 weeks, except for LDL after 11 weeks of storage (change $+9.4\%$). By 27 weeks of storage, triglyceride concentrations had changed by -13.0% in VLDL and by $+13.0\%$ in LDL. The percentage changes in the minor fractions were mostly much larger, but mean absolute changes were again quite small. Although freezing of serum may affect electrophoretic characteristics (4-7) and precipitability (8-9) of lipoproteins with polyanions, cholesterol and triglyceride contents of lipoproteins separated by density-gradient ultracentrifugation appear to be unaffected by prior frozen storage of serum for as long as several months.

We thank our volunteers for the donation of blood, the Netherlands Heart Foundation for funding (grant no. 26.003), Ms. A. Soffers and Mr. F. Schouten for technical and Dr. J. T. Knuiman for editorial assistance.

References

1. Terpstra AHM, Woodward CJH, Sanchez-Muniz FJ. Improved techniques for the separation of serum lipoproteins by density gradient ultracentrifugation: visualization by prestaining and rapid separation of serum lipoproteins from small volumes of serum. *Anal Biochem* 1981;111:149-57.
2. Katan MB, Haar F van der, Kromhout D, et al. Standardization of serum cholesterol assays by use of serum calibrators and direct addition of Liebermann-Burchard reagent. *Clin Chem* 1982;28:683-6.
3. Sullivan DR, Kruijswijk Z, West CE, et al. Determination of serum triglycerides by an accurate enzymatic method not affected by free glycerol. *Clin Chem* 1985;31:1227-8.
4. Milch LJ, Weiner N, Robinson LG. Variability studies of the flotation technique of analytical ultracentrifugation. *J Phys Chem* 1955;59:948-51.
5. DelGatto L, Nichols AV, Lindgren FT. Effects of freeze-thawing and storage on ultracentrifugal properties of human serum lipoproteins. *Proc Soc Exp Biol Med* 1959;101:59-61.
6. Mills GL, Wilkinson PA. Some effects of storage on plasma lipoproteins. *Clin Chim Acta* 1962;7:685-93.
7. Kuchmak M, Taylor L, Olanaky AS. Suitability of frozen and lyophilized reference sera for cholesterol and triglyceride determinations. *Clin Chim Acta* 1982;120:261-71.
8. Bachorik PS, Walker R, Brownell KD, Stunkard AJ, Kwiterovich PO. Determination of HDL-cholesterol levels in stored human plasma samples. *J Lipid Res* 1980;21:608-16.
9. Warnick GR, Albers JJ. High-density-lipoprotein cholesterol (HDL CH) quantitation: effect of plasma storage on heparin-Mn²⁺ supernatant [Abstract]. *Clin Chem* 1979;25:1098.

Effect of High Creatine Content on the Kodak Single-Slide Method for Creatinine, S. F. Sena,¹ D. Syed, and R. B. McComb² (Clin. Chem. Div., Dept. of Pathol. Hartford Hospital, Hartford, CT 06115)

Kodak cautions users of the Ektachem single-slide creatinine method (1) (EKTA/CREA) that high concentrations of creatine in serum may lead to a "Substrate Depletion" flag during the analysis (2). The procedure recommended by Kodak to circumvent this flag is dilution of the specimen and re-analysis. We have encountered this phenomenon on several occasions and herein document these occurrences in two patients.

Case A: This 60-year-old man with a history of angina was admitted to the hospital in cardiac arrest. He was found to have an embolic occlusion and thrombosis in the arterial supply to the right lower extremity for which a thrombectomy and a femoral-popliteal bypass were performed. He developed rhabdomyolysis and acute renal failure with anuria that was treated with hemodialysis. The initial EKTA/CREA analysis (Ektachem 700[®]; Eastman Kodak Co., Rochester, NY 14650) was rejected with a printout of "Out of Range" and "Substrate Depleted". After dilution of the specimen, the creatinine result was 87 mg/L (reference limits 7-17 mg/L) and the serum creatine was 69 mg/L (reference limits 1-4 mg/L). Over the next three weeks the patient's renal function gradually improved, and creatinine concentrations slowly decreased to within adult reference

Fig. 1. Summary of the creatinine results for Patients A and B. The arrows indicate the times that creatine values were determined. The open squares, □, indicate the times when the serum samples were rejected by the EKTACHEM because of substrate depletion. In these cases the creatine concentration at the time of sample rejection is substantially increased. Creatine concentration decreases markedly before creatinine results can be obtained without predilution of sample.

limits (see Figure 1). However, rejection by the Ektachem 700[®] continued until the creatine concentration declined to <31 mg/L.

Case B: This 26-year-old man had a history of intravenous drug abuse and chronic renal failure due to membranoproliferative glomerulonephritis. He was admitted with severe ascites. Lab. studies indicated anemia, marked hypoalbuminemia, and antibodies to human immunodeficiency virus (HIV). The initial serum creatinine by EKTA/CREA was rejected as in Case A; however, after dilution of the specimen, a result of 27 mg/L was obtained. The patient's serum creatine concentration at this time was 136 mg/L. After four days the creatine concentration had declined to 52 mg/L, a concentration that allowed direct measurement of creatinine on the Ektachem 700 (see Figure 1).

Experimental. To better understand concentration relationships between creatine, creatinine, and final creatinine results, we added various concentrations of creatinine and creatine to our "normal" human serum pool (3), then assayed each sample with the Ektachem 700, using EKTA/CREA (generation 1) slides. Any sample rejected as "Out of Range" by the Ektachem was diluted twofold with a 70 g/L solution of bovine albumin in isotonic saline and re-analyzed, as recommended by Kodak (2). The results are summarized in the following tabulation.

Creatinine, mg/L	Creatine concn, mg/L							
	10	20	30	40	50	60	80	100
30	32	31	31	32	32	33	33	34
70	69	68	67	67	68	70	74	75
90	92	90	91	91	94	96	99	98
120	121	121	121	123	125	133	133	137
150	149	148	153	155	166	158	162	165

Results at right are from samples diluted twofold with 70 g/L bovine albumin solution. Each value represents a single analysis.

At all concentrations of creatinine investigated, the threshold concentration of creatine initiating a "Substrate Depletion" flag substantially exceeds the upper reference

¹ Present address: Dept. of Lab. Med., Danbury Hosp., Danbury, CT 06810.

² Author for correspondence.